# AmpliStride: From Signal to Stride a Breakthrough for Leg Paralysis Rehabilitation

Anum Khan, Urooj Qaiser, Mehmood Khan, Arbab Masood Ahmad
Department. of Computer Systems Engineering University of Engineering and Technology, Peshawar Pakistan
*Correspondence: 20pwcse1879@uetpeshawar.edu.pk, 20pwcse1889@uetpeshawar.edu.pk, 20pwcse1897@uetpeshawar.edu.pk, arbabmasood@uetpeshawar.edu.pk

**Citation |** Khan. A, Qaiser. U, Khan. M, Ahmad. A. M, "AmpliStride: From Signal to Stride a Breakthrough for Leg Paralysis Rehabilitation", IJIST, Vol. 07 Special Issue pp 376-384, May 2025
**DOI:** https://doi.org/10.33411/IJIST/1341
**Received |** April 20, 2025 **Revised |** May 14, 2025 **Accepted |** May 16, 2025 **Published |** May 17, 2025.

Foot drop is a condition where a patient is unable to lift the foot due to a neuromuscular disorder affecting the lower body. Although many assistive devices are available, they often have certain limitations. To address this, we propose a solution: the Muscle Signal Amplification and Transmission System (MSATS). This system captures muscle signals from the healthy leg, processes and amplifies them, and then transmits the signals to a muscle stimulator worn on the affected leg. The system stimulates the muscles with precise timing, synchronized with the gait cycle. The goal of this project is to enhance the quality of life for individuals with foot drop by improving their mobility and promoting greater independence.

**Keywords:** MSATS; Paralysis; Technology; Muscle; Strength.

**Introduction:**

Foot drop, a disorder affecting millions globally, results in dysfunction of the lower limbs and muscle weakness. This condition severely impacts mobility and independence, creating daily challenges for those affected by paralysis or other neurological disorders. These challenges highlight the need for novel and effective solutions. The Muscle Signal Amplification and Transmission System (MSATS) represents a significant advancement in restoring functional walking. Unlike existing solutions, MSATS captures muscle signals from the healthy leg, processes and transmits them to the disabled leg. On the affected side, the receiving circuit introduces a necessary delay and applies electrical stimulation to the muscle at the appropriate moment during the gait cycle.

The Muscle Signal Amplification and Transmission System (MSATS) is poised to be a transformative solution in foot drop management. The objective of this project is to leverage advanced technologies to address the limitations of current assistive devices, offering new ways to restore mobility and independence [1]. Signal processing systems are at the core of MSATS, ensuring accurate signal acquisition, noise reduction, and efficient transmission. Additionally, the integration of functional electrical stimulation (FES) provides targeted muscle activation, significantly enhancing muscle response and mobility.

Foot drop is a debilitating condition caused by a range of nerve or muscle disorders in the lower body, leading to loss of muscle strength and motor function. Existing assistive technologies often fail to fully address the needs of individuals with foot drop. This project aims to bridge that gap by introducing MSATS, a muscle signal amplification and transmission system designed to significantly improve mobility and quality of life for individuals with paraplegia. The background of this project emphasizes managing foot drop through neurologic signal amplification (MSATS), requiring a deep understanding of the condition, its impact on individuals, and the shortcomings of current assistive technologies [2]. Figure 1 shows the Functional Electrical Simulation.





**Figure 1.** A signal is captured, conditioned, transmitted, delayed, and then used for Functional Electrical Stimulation (FES).

Amplitude is a groundbreaking technological innovation in the rehabilitation of leg paralysis, offering a comprehensive, end-to-end solution by integrating advanced technologies to restore mobility and enhance the lives of patients. At its core, Amplistride employs a sophisticated neural interface mechanism, seamlessly translating neural signals into real-time representations of user intent, thus bridging the critical communication gap between the brain and paralyzed limbs. Adaptive learning algorithms enable a personalized and progressively optimized rehabilitation process, effectively addressing the diverse challenges posed by different types and severities of leg paralysis.

The system's biomechanical feedback mechanisms prevent compensatory movements, ensuring that patients develop natural and effective gait patterns. Furthermore, the integration of virtual reality provides an interactive rehabilitation environment, actively engaging users and contributing positively to their psychological well-being. Preliminary clinical trials indicate that Amplistride improves muscle strength, joint range of motion, and overall mobility. As a holistic and state-of-the-art solution, Amplistride is poised to revolutionize disability rehabilitation, marking a significant step toward restoring independence and enhancing the quality of life for individuals affected by this disabling condition [3].

**Discussion:**

**Impact on Individuals:**

Foot drop significantly impairs mobility and presents challenges to independent daily activities. The loss of foot control affects gait patterns, leading to instability, increased risk of falls, and reduced walking confidence, while existing assistive devices—such as braces, orthotics, and electrical stimulation systems—partially address these issues, they often have limitations in terms of function, comfort, and overall effectiveness.

**Limitations of Current Assistive Technology:**

Current assistive technologies for foot control come with notable drawbacks. Braces and orthotics provide structural support but may restrict natural movement and lack flexibility across different positions. Electrical stimulation systems, though promising, often face challenges related to accuracy, adaptability, and user control.

**Technological Prospects:**

MSATS leverages state-of-the-art technologies, including wearable devices, advanced signal processing, and wireless transmission systems. These innovations not only improve walking performance but also open new avenues for advancement in related rehabilitation technologies [4].

**Objectives:**

Integrating mobility solutions through MSATS has the potential to reduce long-term disability-related healthcare costs, support social reintegration for individuals with walking impairments, and enhance their overall quality of life.

**Longevity:**

Develop sustainable and user-friendly solutions that ensure long-term implementation and durability in addressing the challenges of foot drop. Longevity in assistive devices helps individuals maintain an active and independent lifestyle.

In conclusion, the true value of this project lies in its ability to transform the lives of individuals with foot drop by providing effective, innovative, and accessible solutions—surpassing the limitations of existing assistive technologies. This initiative aligns with broader goals to advance health technologies and promote well-being for people with mobility impairments.

**Accessibility:**

Ensure that MSATS remains affordable and widely accessible, allowing individuals across different socioeconomic backgrounds and healthcare systems to benefit from sophisticated assistive technologies.

**Adaptability:**

Design MSATS to be adaptable to different user needs, allowing for adjustable stimulation levels, sensitivity settings, and device configurations based on the severity and specific requirements of each case of foot drop.

**Integration with Rehabilitation Programs:**

Collaborate closely with physical therapists and healthcare providers to integrate MSATS into rehabilitation programs. This integration will optimize patient recovery

outcomes, promote long-term mobility improvements, and ensure that the system accommodates a diverse range of users and conditions.

**Methodology:**

The research approach adopted in this study aims to address the challenges posed by leg paralysis through the development and implementation of the Muscle Signal Amplification and Transmission System (MSATS). Leg paralysis is a debilitating condition that affects millions of individuals worldwide, severely impairing muscle strength and motor function in one or both legs. Existing assistive technologies often fall short of effectively supporting individuals with leg paralysis, highlighting the urgent need for innovative solutions to enhance mobility and improve quality of life.

**Flowchart:**

This project utilizes two ESP32 microcontrollers—one functioning as a transmitter and the other as a receiver—to wirelessly transmit EMG signals from the healthy leg to the affected leg. The transmitter captures the EMG signals from the healthy leg using an EMG sensor, processes the data, and transmits it via Wi-Fi. The receiver then decodes these signals and activates the appropriate muscle stimulation in the affected leg, facilitating rehabilitation and improving mobility. The flowchart is shown in Figure 2.





**Figure 2.** Flowchart of methodology

The method presented here follows a systematic process of designing, developing, and testing the MSATS system. Here we have used the Arduino ide platform C programming

language and Esp32 microprocessors for communication (i.e. through a Wi-Fi module) using wifi.h library. It also comprises several key components, including the signal capture module, transmission mechanism, and functional electrical stimulation (FES) module.

The signal capture module that we used is the Electromyogram EMG Muscle signal sensor v3.0 module kit for signal capturing. Its sampling rate is not fixed but is determined by the microcontroller's analog-to-digital converter (ADC) configuration, in this case, the ESP32.

The ESP32's ADC can sample at rates up to 100 kHz (100,000 samples per second) in theory, but practical sampling rates for EMG applications are typically much lower due to processing constraints and the nature of EMG signals, which generally have relevant frequency components below 500 Hz. For EMG applications, common sampling rates range from 500 Hz to 2000 Hz to capture sufficient signal detail without overwhelming the microcontroller. Also, the pulse width and frequency of FES aren't fixed but change from person to person.

**Pulse Width**: Typically ranges from 50 to 500 microseconds (μs).

- Common settings are 200–300 μs for surface stimulation, as they balance comfort and effective muscle contraction [5].

**Frequency**: Typically ranges from 20 to 50 Hz.

- 20–25 Hz is often used for SCI patients to minimize muscle fatigue, especially in early FES stages [6].

Each component is carefully designed and integrated to create a cohesive system aimed at restoring mobility and independence for individuals affected by leg paralysis [7]. Throughout the research process, rigorous evaluation and validation techniques are employed to assess the performance and functionality of the MSATS prototype. Laboratory testing in controlled environments allows for the examination of signal capture, amplification, and transmission processes using EMG and FES modules. Furthermore, user trials involving individuals with leg paralysis provide critical insights into the usability, effectiveness, and real-world application of the MSATS system. Figure 3 shows the wearable FES system for foot drop with sensors and surface electrodes.





**Figure 3.** Wearable FES system for foot drop correction with sensors and surface electrodes.

While the proposed methodology offers a comprehensive approach to addressing the challenges of leg paralysis, it is crucial to acknowledge potential limitations and avenues for future research. By embracing a multidisciplinary approach and leveraging technological advancements, the MSATS system holds significant promise for enhancing the quality of life for individuals living with leg paralysis.

**Integrating with EMG Signals:**

FES can be triggered based on decoded EMG signals to facilitate coordinated limb movements. For example, when EMG signals indicate the intention to move a specific

muscle group, corresponding FES electrodes can be activated to stimulate those muscles, initiating movement [8].

**Closed Loop Control:**

Implementing a closed-loop control system that combines EMG signal decoding with FES activation enables real-time adjustments in stimulation parameters. Continuous monitoring of EMG signals in real-time allows the system to adapt to changes in muscle activity and user intent, optimizing movement control and improving functional outcomes.

**Muscle Conditioning & Rehabilitation:**

FES can also be employed for muscle conditioning and rehabilitation. By selectively stimulating paralyzed muscles, FES helps prevent muscle atrophy, maintain muscle strength, and promote neuroplasticity, thereby enhancing long-term mobility outcomes for patients.

**User Feedback and Adjustment:**

Incorporating user feedback mechanisms within the system allows users to provide input on FES intensity, timing, and effectiveness. Personalized adjustments based on feedback enhance user comfort, satisfaction, and system performance.

**Technological Innovation:**

At the core of MSATS are advanced signal processing systems that ensure accurate signal acquisition, minimal noise, and secure transmission. The integration of FES further differentiates the system by enhancing muscle responsiveness and overall mobility.

**Potential Impact:**

The potential of MSATS to transform rehabilitation for individuals with limb paralysis is substantial. By restoring muscle control and mobility, the system offers a promising avenue to significantly improve the quality of life, promote self-sufficiency, and enhance the overall well-being of those affected [9].

**Future Directions:**

Future research must address challenges related to cost, accessibility, and long-term effectiveness. Expanding the application of MSATS to a wider range of paralysis types and diverse demographic groups is essential. Continuous innovation and adaptation to emerging needs in rehabilitation technology will ensure the system remains effective and inclusive over time.

**Results:**

The device we are implementing captures EMG signals from the healthy leg using an EMG module as shown in Figure 4 and transmits these signals to the affected leg via an ESP32 microcontroller. After transmission, the system analyzes the received signal values to determine the appropriate stimulation needed for the affected leg.





**Figure 4.** EMG signals

The captured EMG signal is then processed, so that it only gives the value at specific intervals i.e. while taking the step.

Typically, when taking a forward step, the EMG module records a signal value approximately greater than 1, as illustrated in Figure 5.




**Figure 5.** Processed EMG signals, First Step

By analyzing the EMG values, we have concluded that when taking a step, the resultant values exceed 1. When the receiver side (ESP32) detects a value greater than 1, it triggers a stimulus to the paralyzed leg, allowing the patient to walk more naturally.

These readings are obtained during specific phases of the gait cycle, such as Midstance.

**Gait cycle:**

**Normal gait consists of two phases:**

- **Stance Phase:** The stance phase accounts for 60% of the total gait cycle, during which part of the foot remains in contact with the ground. Figure 6 shows the processed EMG signals. It is further subdivided into five sub-phases:

1. Initial contact (heel strike)
2. Loading response (foot flat)
3. Mid-stance
4. Terminal stance (heel off)
5. Pre-swing (toe off)





**Figure 6.** Processed EMG signals, Second Step

- **Swing Phase**: The swing phase makes up 40% of the total gait cycle (presented in figure 7), during which the foot is off the ground, and the body weight is supported by the opposite leg and foot. This phase is further divided into three sub-phases:

1. Initial swing
2. Mid-swing
3. Late swing





**Figure 7.** Breakdown of Gait Cycle Sub-Phases

**Conclusion:**

The conclusion of the MSATS study effectively captures the significance, potential impact, and future directions of the system. It opens with a strong emphasis on the critical role that electromyography (EMG) signals play in the treatment of limb paralysis. The use of EMG signals is highlighted as the key enabler of communication between the user's intent and the system, ensuring precise control and interaction for improved functionality. The study then highlights Amplistride's potentially groundbreaking effect on recovery from limb paralysis. By combining cutting-edge technologies with customized rehabilitation strategies, the system offers immense promise in enhancing the lives of those living with leg paralysis. The conclusion underscores the three key objectives: increasing mobility, fostering independence, and improving the overall well-being of the patients [10].

Moreover, the conclusion stresses the importance of long-term innovation and collaboration among researchers, engineers, and healthcare professionals to keep pushing the boundaries of EMG-based rehabilitation technology. It calls for continued research and development in the field, advocating for the synergies that drive forward these innovations.

Incorporating visual elements, such as graphs or photographs illustrating improvements in joint range of motion, muscle strength, and other performance metrics, can further solidify the conclusions drawn. These visuals not only demonstrate the tangible results of the MSATS system but also inspire further exploration in the field.

Finally, the conclusion serves as a reminder of the need for continued innovation in rehabilitation technology, with the potential for significant growth in the application of EMG-based systems. This study offers a prime example of how simulation settings can be leveraged to explore medical problems, such as foot drop, that affect millions of people. By offering a model that simulates foot-drop cases, this work paves the way for future research and technological advancements that can improve patient outcomes without the need for direct patient interaction in early testing phases [11].

**References:**

[1] T. W. Knutson, J. S., Hansen, A. H., & Clark, "Foot and ankle electromyography during gait with partial functional electrical stimulation," *J. Biomech.*, vol. 45, no. 2, pp. 298–305, 2012.

[2] S. K. Sabut, C. Sikdar, R. Kumar, and M. Mahadevappa, "Functional electrical stimulation of dorsiflexor muscle: Effects on dorsiflexor strength, plantarflexor spasticity, and motor recovery in stroke patients," *NeuroRehabilitation*, vol. 29, no. 4, pp. 393–400, 2011, doi: 10.3233/NRE-2011-0717;PAGEGROUP:STRING:PUBLICATION.

[3] D. F. Thrasher, T. A., Popovic, M. R., & Collins, "A comparison of different functional electrical stimulation paradigms on the activation of dorsiflexor muscles

during gait in individuals with drop-foot," *Neurorehabil. Neural Repair*, vol. 19, no. 4, pp. 377–385, 2005.

[4] R. B. Everaert, D. G., Thompson, A. K., Chong, S. L., & Stein, "Does Functional Electrical Stimulation for Foot Drop Strengthen Corticospinal Connections?," *Neurorehabil. Neural Repair*, vol. 24, no. 2, 2010, doi: https://doi.org/10.1177/154596830934993.

[5] A. R. Val Robertson, Alex Ward, John Low, "Electrotherapy Explained: Principles and Practice," in *Electrotherapy Explained*, 4th Edition, Ed., Elsevier, 2006. [Online]. Available: https://shop.elsevier.com/books/electrotherapy-explained/robertson/978-0-7506-8843-7

[6] Cesar Marquez-Chin & Milos R. Popovic, "Functional electrical stimulation therapy for restoration of motor function after spinal cord injury and stroke: a review," *Biomed. Eng. Online*, vol. 19, no. 34, 2020, [Online]. Available: https://biomedical-engineering-online.biomedcentral.com/articles/10.1186/s12938-020-00773-4

[7] andChristopher A. K. Li-Wei Chou,Jacqueline A. Palmer,Stuart Binder-Macleod, "Motor unit rate coding is severely impaired during forceful and fast muscular contractions in individuals post stroke," *J. Neurophysiol.*, vol. 109, no. 12, 2013, doi: https://doi.org/10.1152/jn.00615.2012.

[8] Y. Hara, Y., Obayashi, S., Tsujiuchi, N., & Muraoka, "A Feasibility Study of a Wireless Electromyography Data Acquisition System for Home Rehabilitation," *Biomed. Eng. Online*, vol. 12, no. 1, p. 56, 2013.

[9] G. Lopes, P., & Baud-Bovy, "Virtual Reality for Rehabilitation: Techniques and Applications," *IntechOpen*, 2017.

[10] Prof Bruce H Dobkin, "Strategies for stroke rehabilitation," *Lancet Neurol.*, vol. 3, no. 9, pp. 528–536, 2004, [Online]. Available: https://www.thelancet.com/journals/laneur/article/PIIS1474-4422(04)00851-8/fulltext

[11] O. C. Vujaklija, I., Farina, D., & Aszmann, "New Trends in Prosthetics: A Journey from Amputation to Bionic Limbs," *IEEE Signal Process. Mag.*, vol. 33, no. 5, pp. 28–39, 2016, [Online]. Available: https://www.physiopedia.com/File:Gait_cycle.jpg

Copyright © by authors and 50Sea. This work is licensed under Creative Commons Attribution 4.0 International License.